

Electrical Aerosol Detector

Model 3070A

Standard replacement parts for Model 3070A Electrical Aerosol Detector consist of:

<u>Part Number</u>	<u>Part Description</u>
390065	Aerosol Instrument Manager [®] Software for CPC/EAD 3550/3070A Maintenance Kit: includes (2) 1602230; (1) 1602346; (1) 1602348; (1) 1030319R
1031579	1030319R
1030536	Cyclone 1 micron 3550 & 3070A
1602230	Filter (DFU-BX)
1602346	Filter Wire Core Microfiber
1602348	Carbon Capsule DAU
1030319R	Charger Head Replacement

For pricing information, or to place an order you can 1) send the necessary information to answers@tsi.com or 2) contact your local TSI representative (see contact information below) and speak to one of our customer specialists. Please include your contact information (telephone number and mailing address) with all requests.

Aerosol Instrument Manager[®] is a registered trademark of TSI Incorporated.

TSI Incorporated – 500 Cardigan Road, Shoreview, MN 55126 U.S.A

USA Tel: +1 800 874 2811

UK Tel: +44 149 4 459200

France Tel: +33 491 95 21 90

Germany Tel: +49 241 523030

India Tel: +91 80 41132470

China Tel: +86 10 8260 1595

E-mail: particle@tsi.com

E-mail: tsiuk@tsi.com

E-mail: tsifrance@tsi.com

E-mail: tsigmbh@tsi.com

E-mail: tsi-india@tsi.com

E-mail: tsibeijing@tsi.com

Website: www.tsi.com

Website: www.tsiinc.co.uk

Website: www.tsiinc.fr

Website: www.tsiinc.de



TRUST. SCIENCE. INNOVATION.

Contact your local TSI Distributor or visit our website www.tsi.com for more detailed specifications.